

**Technical
Publications**

5137663-3-800

Revision 1

***Senographe DS Acquisition System
Version ADS 32.04***

**IHE
Integration Statement**

REVISION HISTORY

REFERENCE	DATE	REASON FOR CHANGE
5137663-800 Rev.1	August 5, 2005	Senographe DS milestone M5A.
5137663-2-800 Rev.1	April 3, 2006	Senographe DS milestone M5B - First release.
5137663-3-800 Rev.1	April 11, 2006	Senographe DS milestone M5B - Second release.

INTRODUCTION

OVERVIEW

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on the Internet at: ge.com/ihe

IMPORTANT REMARKS

The use of these IHE Integration Statements, **by itself, is not sufficient to ensure that inter-operation will be successful.** The **user** needs to proceed with caution and ensure that "qualified systems" integration (such as that provided by GE's IT Professional Services) addresses at least four issues:

- **Integration** - The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non-GE systems is the **user's** responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- **Validation** - Testing the complete range of possible interactions between any GE system and non-GE system, before the integration is declared operational, should not be overlooked. Therefore, the **user** should ensure that any non-GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- **Future Evolution** - GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. **In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement.** The **user** should ensure that any non-GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

IHE INTEGRATION STATEMENT

General Electric - IHE Integration Statement		Date
		April 11, 2006
Vendor	Product Name	Version
GE Healthcare	Senographe DS	ADS 32.04
This product implements all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below:		
Integration Profiles Implemented	Actors Implemented	Options Implemented
Scheduled Workflow	Acquisition Modality	Broad Query
Patient Information Reconciliation	Acquisition Modality	None
Internet address for GE Healthcare IHE information: ge.com/ihe		
Links to Standards Conformance Statements for the Implementation:		
HL7	Not applicable	
DICOM	ge.com/dicom See direction #2385079-100	
Links to general information on IHE:		
In North America: www.rsna.org/ihe www.himss.org/ihe	In Europe: www.ihe-europe.org	In Japan: www.jira-net.or.jp/ihe-j

This page is intentionally left blank



GE Healthcare

*GE Healthcare – Americas: Fax 262-544-3384
P.O. Box 414, Milwaukee, Wisconsin, 53201 U.S.A.
Internet – <http://www.gehealthcare.com>*

*GE Healthcare – Europe: Fax 33-1-30-70-95-95
Paris, France
Internet - <http://www.gehealthcareeurope.com>*

*GE Healthcare – Asia
Tokyo, Japan: Fax 81-3-3223-8560
Singapore: Fax 65-291-7006*