



Capsule™-supported connectivity for the Perioperative Care Area

The digital technology revolution in healthcare has changed forever the way that hospitals connect and communicate. Device and network connectivity is essential to administering healthcare in the modern era. Clinicians are facing challenges on all fronts and rising patient acuity requires patient monitoring and therapy devices that connect and capture data with zero compromise to patient care.

When integrating medical devices, vital patient data that is electronically transmitted to the EMR or other third-party application helps save time, reduce charting errors, and allows clinicians more time to spend with their patients.

GE Healthcare and Capsule work together to help ensure connectivity. Medical device connectivity is available using Capsule Medical Device Information System. This solution integrates with your work processes and delivers connectivity across the enterprise in a way that leverages your existing technologies, both inside and outside the GE offering.


GE Healthcare devices currently supported for perioperative care include:

- Aespire*
- Aespire View
- Aestiva*
- Aisys* Carestation*
- Aisys CS²
- Anesthesia Monitor
- Avance* Carestation
- Avance CS²
- Carestation 620
- Carestation 650
- Carestation 650c
- Aware Gateway
- B40 Monitor
- B105 Monitor
- B125 Monitor
- Cardiocap/5
- CARESCAPE* Gateway
- CARESCAPE Monitor B450
- CARESCAPE Monitor B650
- CARESCAPE Monitor B850
- Compact Anesthesia Monitor
- Dash 3000
- Dash 4000
- Dash 5000
- FM Monitor
- Solar* 8000i


Capsule connects with many information systems typically found in the perioperative care environment:

- Amkai
- Dräger
- Epic
- GE Healthcare
- iMDsoft
- iProcedures
- McKesson
- MEDHOST
- Philips
- Picis
- SIS
- SourceMedical





Note: when using Capsule Axons or PCs running Capsule Client, you should use a dedicated virtual network between these devices and their server in order to protect the confidentiality of the collected data.



Capsule cables, DIM and patch cable requirements

Device/Model	Device Serial Cable
B40 Monitor	CBCC402237
B105, B125 Monitors ¹	CBCC402237
Anesthesia Monitor, Compact AM, Cardiocap/5, FM	B1-CFT
CARESCAPE Monitors: B850, B650, B450	B1-CFT and ATEN UC-232A USB
Dash 3000, Dash 4000, Dash 5000	B1-CAQ
Dash 2500	B1-CFB
Aespire, Aespire View, Aisys Carestation, Aisys CS ² Aestiva, Avance Carestation, Avance CS ² , Carestation 600-series (620, 650, 650c)	B1-CFU
DIM	Part Number
Device ID Module	SL-DIM-HW
Patch Cables	Part Number
DIM Patch Cable 5'	B1-CPAT-DIM-5
DIM Patch Cable 14'	B1-CPAT-DIM-14

¹ Requires I/O Cable, GE part number 2099002-015

© 2019 General Electric Company – All rights reserved.

General Electric Company reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation. Contact your GE Representative for the most current information.

GE, GE Monogram, and imagination at work are trademarks of General Electric Company.

GE Healthcare, a division of General Electric Company.

*Trademarks of General Electric Company.

Capsule Medical Device Information System is a product of Capsule Technologies.

Capsule is a trademark of Capsule Technologie SAS.

All other products, company names, brand names, trademarks and logos are the property of their respective owners.

GE Healthcare
P.O. Box 7550
Madison, WI 53707-7550
U.S.A.

DOC1404371 March 2019

