

ROOM PLAN LOCATED ON PAGE 1-166

ITEM	QUANTITY	ITEM DESCRIPTION	WEIGHT	HEAT OUTPUT
①	1	DIGITAL CHEST UNIT	595 lbs	136 btu
②	1	DIGITAL ELEVATING TABLE	992 lbs	372 btu
③	1	SYSTEM CABINET	679 lbs	7442 btu
④	1	RAILS FOR 3M BRIDGE	302 lbs	1501 btu
⑤	1	OPERATORS CONSOLE	71 lbs	3538 btu
⑥	1	XT RADIOGRAPHIC SUSPENSION WITH INBOARD MOUNTING.	654 lbs	1500 btu
***** CUSTOMER/CONTRACTOR SUPPLIED AND INSTALLED ITEMS *****				
⑥0		COUNTER TOP FOR EQUIPMENT- MINIMUM DEPTH 24 IN. AND ADDITIONAL SHELVING MAY BE REQUIRED BELOW COUNTER TOP FOR PC TOWER. PROVIDE GROMMETED OPENINGS AS REQUIRED TO ROUTE CABLES.		
⑥1		MINIMUM DOOR OPENING FOR EQUIPMENT DELIVERY IS 36 IN. W x 66.9 IN. H [914mm x 1700mm], CONTINGENT ON A 96 IN. [2438mm] CORRIDOR WIDTH NOTE: IMAGE PASTE OPTION REQUIRES A 80.9 IN H [2050mm] HIGH OPENING FOR ACCESS.		
⑥2		X-RAY ON WARNING LIGHT - AVAILABLE FROM GE SUPPLY CALL: 800-200-9760 GE CAT. NO. WXIABWW-0F-XIU		
⑥3		DOOR LIMIT SWITCH <NEEDED ONLY IF REQUIRED BY STATE/LOCAL CODES>		
⑥4		CONTROL WALL, 7 FT. HIGH WITH LEAD GLASS VIEWING WINDOW.		
90		X-RAY ROOM WARNING LIGHT CONTROL PANEL REFERENCE JUNCTION POINT 'XRLC' ON SHEET 'E1' FOR DETAILED DESCRIPTION -E4502RL FOR WARNING LIGHT CONTROL ONLY.		
91		EMERGENCY OFF SWITCH <NEEDED ONLY IF REQUIRED BY STATE/LOCAL CODES>		
92		MAIN DISCONNECT, REFERENCE JUNCTION POINT 'A' ON SHEET E1. <16" W x 24" H x 6" D>		

ENVIRONMENT

Ambient Operation Temperature: 59-95° (F)
 Allowable Temperature Change: 10° (C) per hour
 Humidity: 30-80 percent non condensing
 Allowable Humidity Change: <30 percent per hour
 Max - Altitude: 8005 ft. above sea level